

Parameters

Specifications	V800P
Light type	NIR
Wavelength	850nm & 850nm
Basic mode	Yes
Pediatric size	2
Depth detection mode	Yes
Fine mode	Yes
Brightness level	4
Colors	Green, Blue, Red, Violet, White
Radiation value	≤0.6mW/cm ²
Image resolution	854*480pixel
Investigation depth	≤10mm
Optimal imaging distance	210mm±30mm
Net weight	350g
Volume	228*63*62mm
Rechargeable battery	Yes
Standby time	≤4.5h
Charging time	≤3.5h
Stand	Optional



Shenzhen Vivolight Medical Device&Technology Co., Ltd.

Add: 6th Floor, Hi-Tech Building, Block 67, Xingdong Community, Xin'an street, Bao'an District, Shenzhen China
Web: www.veincas.com Tel: +86(0)755 8696 1139 Email: sales@vivo-light.com

Please refer to the user Instruction for contraindications and precautions
Registration certificate: DE / CA05 / MP-238321-2318-00




Safety / Precision / Smart

Projection
Vein
Finder

V800P

- Blood Vessel Assessment
- Venipuncture Optimization
- Blood Vessels Protection




Features

Convenient User Interface




Unique Smart Depth Detection

Provide suggestion of needle entry angle
Red cross aligned with the vein for depth detection
Indication of vein depth by 1/2/3 green blocks




1 green block

Depth: 0-2mm
Needle entry angle: 15°-25°



2 green block

Depth: 2-4mm
Needle entry angle: 25°-30°



3 green block

Depth: 4-6mm
Needle entry angle: ≥30°

“Clear” Fine Mode

Designed for thin vein patients
Reduce the noise to display the veins more clearly



Clinical Application

Indications



Venipuncture



Cosmetic procedures



Vascular procedures

Benefit

Identify
more suitable veins

Increase venipuncture
success rate

Evaluate
vein condition

Improve
patient satisfaction

Time saving
Improve efficiency

Budget saving
Cost-effective

Clinical Recommend

Projection Vein Finder reduces the first venipuncture failure rate by 77.5%, and the infiltration rate by 61.4%.

— Clinical data from a comparative study of 360 cases. The results of this study had been published on Journal of Nursing Administration, September 2015.

The Standard of Care

22.1 To ensure patient safety, the clinician is competent in the use of vascular visualization technology for vascular access device (VAD) insertion. This knowledge includes, but not limited to, appropriate vessels, size, depth, location, and potential complications.

22.2 Vascular visualization technology is used in patients with difficult venous access and/or after failed venipuncture attempts.

22.3 Vascular visualization technology is employed to increase the success with peripheral cannulation and decrease the need for central vascular access device (CVAD) insertion, when other factors do not require a CVAD.

